

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

APR 16 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Datamaster cdm S/N <u>130182</u>	
Name <u>DR. STURKIE</u>	ID# <u>3559</u> Date <u>04-13-2010</u>
A Agency <u>Anchorage Police Dept</u> Phone # <u>786-8900</u>	
Instrument Location <u>Eagle River Substation</u>	
B Alco S/N <u>X124611</u> Target Value <u>.079</u> High Pressure <u>750</u>	
Alco Test Value Average <u>.079</u>	<u>.080</u>
1 st Alco	2 nd Alco
Signature <u>[Signature]</u>	
(OVER)	


(Do Not write in the area below)
I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of April, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130182 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182 ✓

APRIL 13, 2010 ✓ *COB 4/26/10*

OPERATOR'S NAME:

STURKIE

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: ANCL

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124611 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.67 in		
ALCO TARGET	.078	04:39
BLANK TEST	.000	04:39
INTERNAL STANDARD	VERIFIED	04:40
ALCO TV 29.67 in	.079 ✓	04:40
BLANK TEST	.000	04:41
SUBJECT SAMPLE	.000	04:41
BLANK TEST	.000	04:42
ALCO TV 29.67 in	.080 ✓	04:42
BLANK TEST	.000	04:43

COB 4/26/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182 ✓

APRIL 13, 2010 ✓
TIME 04:35

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.67 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz
pqrstuvwxyz{|}~△